

# シリンジ検量証明書

**Trajanのシリンジは検量証明書が添付可能です。**  
シリンジの信頼性を確かなものにします。

Trajan ではシリンジの注文時に指定していただければ、シリンジの出荷前に高精度な測定器を使用し、検量を行い、検量証明書を添付いたします（検量費用が別途必要となります）。

固有のシリアルナンバーを、証明書とシリンジボディに印字しますので、シリンジの識別が可能となります。検量証明書付きシリンジをご注文の際は、シリンジのパーツナンバーの後ろに「CAL」と付けてご注文下さい。

例：固定針型 10  $\mu$ L 標準シリンジ 10F（PartNo.002000）の場合は「002000CAL」でご注文下さい。

※マルチパックシリンジの場合は、「マルチパックシリンジ価格 + (入数  $\times$  ¥18,000)」となりますのでご注意ください。

詳細	Part no.	価格
検量証明書	000000 + CAL	シリンジ価格 + 入数 $\times$ ¥18,000

シリンジパーツナンバー

シリンジコード(品名)

シリアルナンバー


プランジャーチップ外径  
(ガスタイトシリンジではガラスパレル内径)

目盛長さ

フルスケールボリューム

針ボリューム

日付



This certificate of calibration provides traceable displacement volume accuracy for the enclosed syringe. It should be noted that the precision of dispensed volume will depend on operating conditions and operator technique.

Volume accuracy has been measured using equipment traceable to International Standards.


A copy of Trajan Scientific and Medical's method of calibration is available on request.

Part number	007250CAL	
Code	500R-GT 500UL SYRINGE	
Serial number	R126	
Plunger diameter at tip (barrel diameter for gas tight syringes)	3.247 mm	$\pm$ 0.001 mm
Scale length	60.010 mm	$\pm$ 0.01 mm
Full syringe capacity	496.912 $\mu$ L	$\pm$ 0.389 $\mu$ L
Needle volume (nominal)	1.649 $\mu$ L	

Date 01 / 01 / 20--      Signature (Quality Control)

Certificate of calibration

担当者サイン



www.trajanscimed.com